



Cascade Engineering Services, Inc.

12026 115th Ave NE, Suite 102, Kirkland, WA 98034

T.425.895.8617, F.425.702.9358



CERTIFICATE #: 00121645

CERTIFICATE OF CALIBRATION

STANDARD CALIBRATION

CLYDE HILL POLICE DEPARTMENT

9605 NE 24TH ST. ,

This certifies that the instrument listed herein was calibrated by Cascade Engineering Services' Calibration Laboratory, which is fully accredited in accordance with the recognized International Standards ISO/IEC 17025:2005 General Requirements for the Competence of Testing and Calibration Laboratories. Cascade Engineering Services' Calibration Laboratory also meets the requirements of ANSI/NCSL Z540-1-1994 and any additional program requirements in the field of calibration. Standards used to perform this calibration are certified by or traceable to NIST, natural physical constants, consensus standards or derived by the ratio type of calibrations. All calibrations are performed to manufacturer's specifications unless otherwise noted. Standard Calibration, while still traceable, does not meet all requirements for an Accredited Calibration per ISO/IEC 17025:2005, that is "As Found" data for equipment in tolerance and Measurement Uncertainties are not recorded. This certificate shall not be reproduced, except in full, without prior written approval of the laboratory.

DESCRIPTION:	SPEED MEASURING DEVICE		
ASSET #:	FF12050	SERIAL NUMBER:	FF12050
MANUFACTURER:	KUSTOM	MODEL NUMBER:	FALCON
DEPARTMENT:	N/A	PATROL CAR #:	
ENVIRONMENT:	71.0 °F/38.7 %RH	BASIC ACCURACY:	FUNCTIONAL TEST ONLY
CAL INTERVAL:	12 MONTHS	DUE DATE:	October 29, 2015

EQUIPMENT CONDITION AS RECEIVED

Initial testing found this equipment to be in the condition stated in the certification notes below.

EQUIPMENT CONDITION AS DELIVERED

At the completion of the calibration, measured values were "IN TOLERANCE", as defined by the basic accuracy stated above.

TUNNING FORK(S) SUPPLIED WITH THIS DEVICE

DESCRIPTION	SERIAL NUMBER	RATED SPEED	FREQUENCY
TUNNING FORK ONE	15424	50 M.P.H.	3652 Hz
TUNNING FORK TWO	N/A		

Antenna 1 SN: HANDHELD, Frequency: 24.165 GHz

STANDARD(S) USED FOR CERTIFICATION

I.D.	MODEL	MANUFACTURER	DESCRIPTION	DUE DATE
MET1231	VOCAR HR	DB INNOVATIONS	HAND HELD RADAR CERTIFICATION SYSTEM	01/13/2015
MET1232	VOCAR HR WAND	DB INNOVATIONS	VOCAR HR WAND	01/13/2015

CERTIFICATION NOTES

REPLACED BROKEN ANT. LENS.
BENCH TEST OK.

I certify (or declare) under penalty of perjury under the laws of the State of Washington that the above information is true and correct

PERFORMED BY

Charlie M. Brown
CALIBRATION TECHNICIAN: CHARLIE BROWN

LOCATION: Kirkland, Wa

CALIBRATION DATE: Wednesday, October 29, 2014

THIS LABORATORY IS A2LA ACCREDITED TO ISO/IEC 17025:2005 (GENERAL REQUIREMENTS FOR THE COMPETENCE OF TESTING AND CALIBRATION LABORATORIES), CERTIFICATE #: 2580.01

This agency, CLYDE HILL POLICE DEPARTMENT currently utilizes the following Laser SMD(s):


KUSTOM Manufacturer's the following SMD(s)

ID/Serial Number	Model Number	Antenna 1 S/N	Antenna 2 S/N	T.F. 1 S/N	T.F. 2 S/N	Cal. Date	Cal. Interval	Due Date
PL19672	PRO LASER III	N/A	N/A	N/A	N/A	05/27/2014	12 MONTHS	05/27/2015

This agency, CLYDE HILL POLICE DEPARTMENT currently utilizes the following Doppler SMD(s):

KUSTOM Manufacturer's the following SMD(s)

ID/Serial Number	Model Number	Antenna 1 S/N	Antenna 2 S/N	T.F. 1 S/N	T.F. 2 S/N	Cal. Date	Cal. Interval	Due Date
BB6583	KR-10SP	8308	N/A	30613	29227	05/27/2014	12 MONTHS	05/27/2015
EE18185	KR-10SP	CC17576	N/A	45633	43464	08/13/2014	12 MONTHS	06/13/2015
FF12050	FALCON	HANDHELD	N/A	15424	N/A	10/29/2014	12 MONTHS	10/29/2015
KK12693	TROOPER	CC13284	N/A	32536	31231	07/29/2014	12 MONTHS	07/29/2015
KK2711	TROOPER	CC13283	N/A	55529	51537	10/16/2014	12 MONTHS	10/16/2015
T3306/T3308K	TALON	HANDHELD	N/A	40141	N/A	08/07/2014	12 MONTHS	08/07/2015



Certified by: Charles N. Brown
Place: Redmond, WA

October 29, 2014
Date