



Cascade Engineering Services, Inc.

6640 185th Ave NE, Redmond, WA 98052
T.425.895.8617, F.425.702.9358

FILED

AUG - 2 2022

KIRKLAND MUNICIPAL COURT



CERTIFICATE #: 00245254

CERTIFICATE OF CALIBRATION

STANDARD CALIBRATION

CLYDE HILL POLICE DEPARTMENT

9605 NE 24TH ST. CLYDE HILL, WA 98004

This certifies that the instrument listed herein was calibrated by Cascade Engineering Services' Calibration Laboratory, which is fully accredited in accordance with the recognized International Standards ISO/IEC 17025:2017 General Requirements for the Competence of Testing and Calibration Laboratories.

DESCRIPTION:

SPEED MEASURING DEVICE

Table with 2 columns: Description and Value. Rows include ASSET #, MANUFACTURER, DEPARTMENT, ENVIRONMENT, CAL INTERVAL, SERIAL NUMBER, MODEL NUMBER, PATROL CAR #, BASIC ACCURACY, and DUE DATE.

EQUIPMENT CONDITION AS RECEIVED

Initial testing found this equipment to be "IN TOLERANCE", as defined by the basic accuracy stated above.

EQUIPMENT CONDITION AS DELIVERED

At the completion of the calibration, measured values were "IN TOLERANCE", as defined by the basic accuracy stated above.

LASER SYSTEM OUTPUTS

PULSE REFERENCE FREQUENCY 199.9940 Hz OPTICAL POWER OUTPUT: 184.7 uW

STANDARD(S) USED FOR CERTIFICATION

Table with 5 columns: I.D., MODEL, MANUFACTURER, DESCRIPTION, DUE DATE. Lists standards like NOVA-DISPLAY, PD300-SH, 7024707, 7005320.

PROCEDURE(S) USED FOR CERTIFICATION

Table with 4 columns: DOCUMENT ID, DESCRIPTION, REV, REV DATE. Lists CP-SMD-001 RADAR AND LIDAR CALIBRATION PROCEDURE.

CERTIFICATION NOTES

I certify (or declare) under penalty of perjury under the laws of the State of Washington that the above information is true and correct

PERFORMED BY [Signature] METROLOGIST JOHN GRAY

LOCATION: Redmond, WA CALIBRATION DATE: Tuesday, August 02, 2022

THIS DOCUMENT IS MAINTAINED AS A PUBLIC RECORD IN ACCORDANCE WITH RCW 5.44

METROLOGY LAB SHIPPER

SHIPPING REMARKS:

SHIP TO:

COMPANY: CLYDE HILL POLICE DEPARTMENT
CONTACT: ELLIE SILVEIRA
ADDRESS: 9605 NE 24TH ST.
 CLYDE HILL, WA 98004
PHONE: (425) 454-7187

FROM:

VENDOR: CES Calibration Services
CONTACT: Cascade Engineering Services, Inc.
ADDRESS: 6640 185TH AVE NE
 Redmond, WA 98052
PHONE: 425.895.8617 ext. 707

PO Number: INVOICE

EQUIPMENT INCLUDED:

	MANUFACTURER		DESCRIPTION	DEPARTMENT
1	PL19672	KUSTOM SIGNALS, INC. PRO LASER III	SPEED MEASURING DEVICE	N/A

CUSTOMER SIGNATURE: _____

DATE RECEIVED _____

SHIPPED BY: _____

JOHN GRAY

SHIPPING DATE: August 2, 2022



Cascade Engineering Services, Inc.

6640 185th Ave. NE, Redmond, WA 98052
T.425.895.8617, F.425.702.9358



CERTIFICATION CONCERNING DESIGN AND CONSTRUCTION OF
ELECTRONIC SPEED MEASURING DEVICES

IRLJ RULE 6.6 EFFECTIVE 1/3/2006

I, Clifford W. Nahm, do certify under penalty of perjury, under the laws of the state of Washington as follows:

I am employed with Cascade Engineering Services, Inc. (CES) Metrology and Electronics Repair Services, as the Metrology Lab Manager. I have been employed in such a capacity since 1996. Part of my duties includes supervising the maintenance and repair of all doppler radar and laser speed measuring devices (SMD's) used by CLYDE HILL POLICE DEPARTMENT.

All SMD's currently used by CLYDE HILL POLICE DEPARTMENT are listed in Exhibit "A".

I maintain the following qualifications with respect to SMD(s): United States Navy trained electronics technician specializing in Precision Measurement Electronics (PMEL) C School & Systems/Communications/Radar A School. I have commercial experience in electronics and in the repair and calibration of Doppler and Lidar SMD's dating back to 2006. I am experienced and competent in the principles and fundamental requirements of calibration from DC to Microwave frequencies.

The CES laboratory maintains manuals for all of the SMD's listed in Exhibit "A". I am personally familiar with those manuals and how each of the SMD's are designed and operated. On the date indicated in Exhibit "A" testing of the SMD's was performed using CES procedures under the direction of an authorized SMD expert. The results were evaluated and certified to meet or exceed existing performance standards and entered into the CES certification management database. CES laboratory maintains a testing and certification program that requires each SMD to be tested and certified for accuracy at least once every two years.

The CES laboratory tests all Doppler SMD's used by CLYDE HILL POLICE DEPARTMENT, as recommended by the manufacturer, as follows: The Vocar HR, handheld Radar certification system is used to simulate speeds at 5 mph increments from 20 mph to 140 mph to verify accuracy in stationary and moving mode. Measurements are taken of the SMD transmit frequency, antenna/receiver sensitivity and any accompanying tuning forks are also tested for accuracy. All other operational functions of the SMD system are then tested for proper performance.

The Laser SMD's transmit a series of highly focused light wave pulses each time the trigger is pulled and utilizes two laws of physics; time and distance (i.e. 3.5 feet in diameter at 1000 feet). Since the speed of light is a known fixed value, the distance of the target is determined by calculating how long it takes for the signal to travel to the target and back. This series of measurements allows the SMD to calculate the speed of the target by measuring the distance the signal took to travel to the target and back. The displayed speed is accurate to within plus (+) or minus (-) one (1) mile per hour.

The CES laboratory tests all Laser / Lidar SMD(s) used by CLYDE HILL POLICE DEPARTMENT, as recommended by the manufacturer, as follows: The Laser Speed Measurement Simulator (LSMS) is utilized to simulate a moving target. This is accomplished by detecting the optical output pulses of the laser device and generating artificial return pulses. Different speed values and ranges are simulated by varying the time delays between the input pulses and the return pulses. The LSMS consists of a Digital Delay Generator (DDG), and an optical interface unit. The DDG produces precise time delays. The optical interface unit converts the optical energy of the laser instrument into electrical signals which are supplied to the DDG. The optical interface unit also converts the electrical signals received from the DDG into optical energy which is then transmitted to the Lidar. The Lidar's output power is tested using an Ophir Nova Display, with a PD300-SH power head.

On the date indicated in Exhibit "A", each SMD was tested by a trained technician listed therein and under my direction. All Technicians listed on Exhibit "A" received training in the proper use and operation of SMD test equipment and performance testing procedures used to test Laser and Doppler SMDs. After successfully completing training the technician is certified by myself and receives authorization allowing them to enter the results from the tests into the certificate management database. Individual Performance and Certification tests are entered into the certificate management database under the penalty of perjury by entering an authorized user id and password to authenticate it.

Exhibit "A"

This agency, CLYDE HILL POLICE DEPARTMENT currently utilizes the following Laser SMD(s):

KUSTOM SIGNALS, INC. manufacture's the following SMD(s):

I.D./Serial Number	Model Number	Antenna 1 S/N	Antenna 2 S/N	T.F. 1 S/N	T.F. 2 S/N	Cal. Date	Cal. Interval	Due Date	Technician
PL19672	PRO LASER III	N/A	N/A	N/A	N/A	08/02/2022	12 MONTHS	08/02/2023	JOHN R GRAY

This agency, CLYDE HILL POLICE DEPARTMENT currently utilizes the following Doppler SMD(s):

KUSTOM SIGNALS, INC. manufacture's the following SMD(s):

I.D./Serial Number	Model Number	Antenna 1 S/N	Antenna 2 S/N	T.F. 1 S/N	T.F. 2 S/N	Cal. Date	Cal. Interval	Due Date	Technician
FH12675	FALCON	HANDHELD	N/A	33281	31332	06/07/2022	12 MONTHS	06/07/2023	JOHN R GRAY
KK37897	TROOPER	CC37517	N/A	3303	3136	04/07/2022	12 MONTHS	04/07/2023	JOHN R GRAY
KK12693	TROOPER	CC13284	N/A	32536	31231	04/07/2022	12 MONTHS	04/07/2023	JOHN R GRAY

MPH INDUSTRIES manufacture's the following SMD(s):

I.D./Serial Number	Model Number	Antenna 1 S/N	Antenna 2 S/N	T.F. 1 S/N	T.F. 2 S/N	Cal. Date	Cal. Interval	Due Date	Technician
PYT124000537	PYTHON III	PYT831014894	PYT831014893	64737	64321	07/12/2022	12 MONTHS	07/12/2023	JOHN R GRAY
PYT124000663	PYTHON III	PYT831015736	PYT831015737	70621	69972	06/07/2022	12 MONTHS	06/07/2023	JOHN R GRAY
PYT124000196	PYTHON III	PYT831012980	PYT831012981	867661	N/A	04/07/2022	12 MONTHS	04/07/2023	JOHN R GRAY

Based upon my education, training, and experience and my knowledge of the SMD's listed above, it is my opinion that each of these electronic pieces of equipment is so designed and constructed as to accurately employ the Doppler effect in such a manner that it will give accurate measurements of the speed of motor vehicles when properly calibrated and operated by a trained operator or, in the case of the laser SMDs, each of these pieces of equipment is so designed and constructed as to accurately employ measurement techniques based on the velocity of light in such a manner that it will give accurate measurements of the speed of motor vehicles when properly calibrated and operated by a trained operator.

Exhibit "A" derives information from the certificate management database. See Exhibit "A" for details about individual SMD certifications.


State of Washington
County of King


Signed or attested before me on

8-2-2022 by Clifford W. Nahm

I have satisfactory evidence that the person described in this document:

- (a) is personally known to me; OR (b) is identified upon oath or affirmation of credible witness personally known to me; OR
- (c) is identified on the basis of Identification documents.


Certified by: Clifford W. Nahm
Place: Redmond, WA


William Quoc Ang
Notary Public in and for the State of Washington,
Residing in Seattle, WA
My appointment expires January 29, 2026

