



Cascade Engineering Services, Inc.

6640 185th Ave. NE, Redmond, WA 98052
T.425.895.8617, F.425.702.9358



**CERTIFICATION CONCERNING DESIGN AND CONSTRUCTION OF
ELECTRONIC SPEED MEASURING DEVICES**

IRLJ RULE 6.6 EFFECTIVE 1/3/2006

I, Nicolas T. Mowry, do certify under penalty of perjury, under the laws of the state of Washington as follows:

I am employed with Cascade Engineering Services, Inc. (CES) Metrology and Electronics Repair Services, as a Metrology Technician. I have been employed in such a capacity since 2016. Part of my duties includes supervising the maintenance and repair of all electronic and laser speed measuring devices (SMD's) used by CLYDE HILL POLICE DEPARTMENT.

All SMD's currently used by CLYDE HILL POLICE DEPARTMENT are listed in Exhibit "A".

I maintain the following qualifications with respect to SMD(s): I have commercial experience in electronics and in the repair and calibration of Doppler and Lidar SMD's since the beginning of 2016. I have a Bachelors of Science degree in physics from the University of Washington. I am experienced and competent in the principles and fundamental requirements of calibration from DC to Microwave frequencies.

The CES laboratory maintains manuals for all of the SMD's listed in Exhibit "A". I am personally familiar with those manuals and how each of the SMD's are designed and operated. On the date indicated in Exhibit "A" testing of the SMD's was performed using CES procedures under the direction of an authorized SMD expert. The results were evaluated and certified to meet or exceed existing performance standards and entered into the CES certification management database. CES laboratory maintains a testing and certification program that requires each SMD to be tested and certified for accuracy at least once every two years.

The CES laboratory tests all Doppler SMD's used by CLYDE HILL POLICE DEPARTMENT, as recommended by the manufacturer, as follows: The Vocar HR, handheld Radar certification system is used to simulate speeds at 5 mph increments from 20 mph to 140 mph to verify accuracy in stationary and moving mode. Measurements are taken of the SMD transmit frequency, antenna/receiver sensitivity and any accompanying tuning forks are also tested for accuracy. All other operational functions of the SMD system are then tested for proper performance.

The Laser SMD's transmit a series of highly focused light wave pulses each time the trigger is pulled and utilizes two laws of physics; time and distance (i.e. 3.5 feet in diameter at 1000 feet). Since the speed of light is a known fixed value, the distance of the target is determined by calculating how long it takes for the signal to travel to the target and back. This series of measurements allows the SMD to calculate the speed of the target by measuring the distance the signal took to travel to the target and back. The displayed speed is accurate to within plus (+) or minus (-) one (1) mile per hour.

The CES laboratory tests all Laser / Lidar SMD(s) used by CLYDE HILL POLICE DEPARTMENT, as recommended by the manufacturer, as follows: The Laser Speed Measurement Simulator (LSMS) is utilized to simulate a moving target. This is accomplished by detecting the optical output pulses of the laser device and generating artificial return pulses. Different speed values and ranges are simulated by varying the time delays between the input pulses and the return pulses. The LSMS consists of a Digital Delay Generator (DDG), and an optical interface unit. The DDG produces precise time delays. The optical interface unit converts the optical energy of the laser instrument into electrical signals which are supplied to the DDG. The optical interface unit also converts the electrical signals received from the DDG into optical energy which is then transmitted to the Lidar. The Lidar's output power is tested using an Ophir Nova Display, with a PD300-SH power head.

On the date indicated in Exhibit "A", each SMD was tested by a trained technician listed therein and under my direction. All Technicians listed on Exhibit "A" received training in the proper use and operation of SMD test equipment and performance testing procedures used to test Laser and Doppler SMDs. After successfully completing training the technician is certified by myself and receives authorization allowing them to enter the results from the tests into the certificate management database. Individual Performance and Certification tests are entered into the certificate management database under the penalty of perjury by entering an authorized user id and password to authenticate it.

FILED

AUG 14 2020

**KIRKLAND
MUNICIPAL COURT**

Exhibit "A"

This agency, CLYDE HILL POLICE DEPARTMENT currently utilizes the following Laser SMD(s):

KUSTOM SIGNALS, INC. manufacturer's the following SMD(s):

I.D./Serial Number	Model Number	Antenna 1 S/N	Antenna 2 S/N	T.F. 1 S/N	T.F. 2 S/N	Cal. Date	Cal. Interval	Due Date	Technician
PL19672	PRO LASER III	N/A	N/A	N/A	N/A	08/13/2020	12 MONTHS	08/13/2021	NICOLAS T MOWRY

This agency, CLYDE HILL POLICE DEPARTMENT currently utilizes the following Doppler SMD(s):

KUSTOM SIGNALS, INC. manufacturer's the following SMD(s):

I.D./Serial Number	Model Number	Antenna 1 S/N	Antenna 2 S/N	T.F. 1 S/N	T.F. 2 S/N	Cal. Date	Cal. Interval	Due Date	Technician
FH12675	FALCON	HANDHELD	N/A	33281	31332	06/22/2020	12 MONTHS	06/22/2021	NICOLAS T MOWRY
KK12693	TROOPER	CC13284	N/A	32536	31231	04/29/2020	12 MONTHS	04/29/2021	NICOLAS T MOWRY
KK37897	TROOPER	CC37517	N/A	3303	3136	04/29/2020	12 MONTHS	04/29/2021	NICOLAS T MOWRY
KK2711	TROOPER	CC13283	N/A	55529	51537	11/26/2019	12 MONTHS	11/26/2020	NICOLAS T MOWRY

MPH INDUSTRIES manufacturer's the following SMD(s):

I.D./Serial Number	Model Number	Antenna 1 S/N	Antenna 2 S/N	T.F. 1 S/N	T.F. 2 S/N	Cal. Date	Cal. Interval	Due Date	Technician
PYT124000196	PYTHON III	PYT831012980	PYT831012981	867661	50183	04/29/2020	12 MONTHS	04/29/2021	NICOLAS T MOWRY

Based upon my education, training, and experience and my knowledge of the SMD's listed above, it is my opinion that each of these electronic pieces of equipment is so designed and constructed as to accurately employ the Doppler effect in such a manner that it will give accurate measurements of the speed of motor vehicles when properly calibrated and operated by a trained operator or, in the case of the laser SMDs, each of these pieces of equipment is so designed and constructed as to accurately employ measurement techniques based on the velocity of light in such a manner that it will give accurate measurements of the speed of motor vehicles when properly calibrated and operated by a trained operator.

Exhibit "A" derives information from the certificate management database. See Exhibit "A" for details about individual SMD certifications.

State of Washington

County of King

Signed or attested before me on

8-14-2020

by Nicolas T. Mowry

I have satisfactory evidence that the person described in this document:

(a) is personally known to me; OR (b) is identified upon oath or affirmation of credible witness personally known to me; OR (c) is identified on the basis of identification documents.



William Quoc Ang

William Quoc Ang
Notary Public in and for the State of Washington,
Residing in Seattle, WA
My appointment expires January 29, 2022

Nicolas T. Mowry

Certified by: Nicolas T. Mowry
Place: Redmond, WA