



Cascade Engineering Services, Inc.

6640 185th Ave NE, Redmond, WA 98052

T.425.895.8617, F.425.702.9358

FILED

MAR - 9 2020

KIRKLAND
MUNICIPAL COURT



CERTIFICATE #: 00204174

CERTIFICATE OF CALIBRATION

STANDARD CALIBRATION

CLYDE HILL POLICE DEPARTMENT

9605 NE 24TH ST. CLYDE HILL, WA 98004

This certifies that the instrument listed herein was calibrated by Cascade Engineering Services' Calibration Laboratory, which is fully accredited in accordance with the recognized International Standards ISO/IEC 17025:2005 General Requirements for the Competence of Testing and Calibration Laboratories. Cascade Engineering Services' Calibration Laboratory also meets the requirements of ANSI/NCSL Z540-1-1994 and any additional program requirements in the field of calibration. Standards used to perform this calibration are certified by or traceable to NIST, natural physical constants, consensus standards or derived by the ratio type of calibrations. All calibrations are performed to manufacturer's specifications unless otherwise noted. Standard Calibration, while still traceable, does not meet all requirements for an Accredited Calibration per ISO/IEC 17025:2005, that is "As Found" data for equipment in tolerance and Measurement Uncertainties are not recorded. This certificate shall not be reproduced, except in full, without prior written approval of the laboratory.

DESCRIPTION:

SPEED MEASURING DEVICE

ASSET #:	PYT124000196	SERIAL NUMBER:	PYT124000196
MANUFACTURER:	MPH INDUSTRIES	MODEL NUMBER:	PYTHON III
DEPARTMENT:	N/A	PATROL CAR #:	N/A
ENVIRONMENT:	67.8 °F/43.7 %RH	BASIC ACCURACY:	REFERENCE MFRS SPECIFICATIONS
CAL INTERVAL:	12 MONTHS	DUE DATE:	May 23, 2020

EQUIPMENT CONDITION AS RECEIVED

Initial testing found this equipment to be "IN TOLERANCE", as defined by the basic accuracy stated above.

EQUIPMENT CONDITION AS DELIVERED

At the completion of the calibration, measured values were "IN TOLERANCE", as defined by the basic accuracy stated above.

TUNING FORK(S) SUPPLIED WITH THIS DEVICE

DESCRIPTION	SERIAL NUMBER	RATED SPEED	FREQUENCY
TUNING FORK ONE	867661	35 M.P.H.	2520 Hz
TUNING FORK TWO	50183	65 M.P.H.	4671 Hz

Antenna 1 SN: PYT831012980, Frequency: 24.154 GHz

Antenna 2 SN: PYT831012981, Frequency: 24.155 GHz

STANDARD(S) USED FOR CERTIFICATION

I.D.	MODEL	MANUFACTURER	DESCRIPTION	DUE DATE
MET1231	VOCAR HR	DB INNOVATIONS	HAND HELD RADAR CERTIFICATION SYSTEM	08/24/2019
MET1232	VOCAR HR WAND	DB INNOVATIONS	VOCAR HR WAND	08/24/2019

PROCEDURE(S) USED FOR CERTIFICATION

DOCUMENT ID	DESCRIPTION	REV	REV DATE
SMD101	DOPPLER RADAR / LIDAR CALIBRATION PROCEDURE	A	06/01/2006

CERTIFICATION NOTES

I certify (or declare) under penalty of perjury under the laws of the State of Washington that the above information is true and correct

PERFORMED BY

Nicolas Mowry
SENIOR METROLOGIST NICOLAS MOWRY

LOCATION: Redmond, WA

CALIBRATION DATE: Thursday, May 23, 2019

THIS LABORATORY IS A2LA ACCREDITED TO ISO/IEC 17025:2005 (GENERAL REQUIREMENTS FOR THE COMPETENCE OF TESTING AND CALIBRATION LABORATORIES) CERTIFICATE #: 2560.01
F540-1.3/REVISION L
JUNE-2011

Page 1 of 1

THIS DOCUMENT IS MAINTAINED
AS A PUBLIC RECORD IN
ACCORDANCE WITH RCW 5.44