

FILED

APR 8 2020

KIRKLAND
MUNICIPAL COURT

CERTIFICATE OF ACCURACY

 **KUSTOM SIGNALS, INC.**
 1010 WEST CHESTNUT, PO BOX 947 CHANUTE, KS 66720-0947

This is to certify that on the 17 day of March,
2020 the below instrument met the manufacturer's specifications.
 Manufacturer: Kustom Signals, Inc.
 Indicator type: RADAR RP-1
 Serial Number: RP 25359
 MPH KM/H KNOTS OTHER

The instruments used to certify the above equipment have been calibrated within the previous year and calibration is traceable to the National Institute of Standards and Technology.

Signed: [Signature] Title: Technician
 Subscribed and sworn to me this 17 day of March, 2020

NOTARY PUBLIC [Signature]



006-0725-00 Rev. 2

THIS DOCUMENT IS MAINTAINED
 AS A PUBLIC RECORD IN
 ACCORDANCE WITH RCW 5.44

FILED

APR 8 2020

KIRKLAND
MUNICIPAL COURT

**Kustom Signals, Inc.
Tuning Fork Certification of Accuracy**

This is to certify that

Kustom K-Band Tuning Fork SN 50746 has been tested and found to oscillate at 2544 HERTZ at 21 °C. When used with a Kustom K-Band Doppler traffic radar operating at 24150 MHz, will cause a calibration signal of 35 MPH.

(Temperature correction factor: -0.02 MPH/°C)

DATE 3/14/2020 Certified by

Ayalee Bright

The instrument used to certify the frequency of the above tuning fork has been calibrated within the previous year and is traceable to the National Institute of Standards and Technology.



1010 W. CHESTNUT CHANUTE, KANSAS 66720



006-0955-00

FILED

APR 8 2020

KIRKLAND
MUNICIPAL COURT

**Kustom Signals, Inc.
Tuning Fork Certification of Accuracy**

This is to certify that

Kustom K-Band Tuning Fork SN 50815 has been tested and found to oscillate at 4734 HERTZ at 21 °C. When used with a Kustom K-Band Doppler traffic radar operating at 24150 MHz, will cause a calibration signal of 65 MPH.

(Temperature correction factor: -0.02 MPH/°C)

DATE 3/14/2020 Certified by



The instrument used to certify the frequency of the above tuning fork has been calibrated within the previous year and is traceable to the National Institute of Standards and Technology.



1010 W. CHESTNUT CHANUTE, KANSAS 66720



006-0955-00

THIS DOCUMENT IS MAINTAINED
AS A PUBLIC RECORD IN
ACCORDANCE WITH RCW 5.41